

CERTIFICATE OF ANALYSIS

PRODUCT :	CLED Agar
PRODUCT CODE NO.:	PP006-20

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/ STANDARD VALUES
Colour	Green colour
Clarity	Slightly opalescent
Appearance	Sterile CLED Agar w/Bromo Thymol Blue in 85mm disposable plates
Sterility after 24hrs	Sterile

QUALITY CONTROL TESTS: MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed after (18-24) hours at (35-37)oC .

Orgm Desc	Inoculum(CFU)	Growth	Recovery	Colour of colony
Enterococcus faecalis ATCC 29212	50-100	good-luxuriant	>=70%	slight yellowish or greenish
Escherichia coli ATCC 25922	50-100	good-luxuriant	>=70%	yellow, opaque, centre slightly deeper yellow
Klebsiella pneumoniae ATCC 13883	50-100	good-luxuriant	>=70%	yellow to whitish blue
Proteus vulgaris ATCC 13315	50-100	good-luxuriant	>=70%	blue
Salmonella typhi ATCC 6539	50-100	good-luxuriant	>=70%	bluish
Staphylococcus aureus ATCC 25923	50-100	good-luxuriant	>=70%	deep yellow

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.