

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT: D.T.M. Agar (Dermatophyte Test Agar Base) Plate**

**PRODUCT CODE NO: PP007-20**

### QUALITY CONTROL TESTS: DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Orange Red coloured medium
APPEARANCE	Dermatophyte Test Agar in 90X15mm disposable plates.
CLARITY	Slightly opalescent
STERILITY AFTER 48 HRS	Sterile
PH	5.5 ± 0.2

### QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE:

Cultural characteristics observed after incubation at 22 - 28°C for up to 6 days.

Organism Description	Inoculum(CFU)	Growth	COLOR OF COLONY
Trichophyton mentagrophytes 9533	[50-100]	[Good]	[Pink-red]
Candida albicans 10231	[50-100]	[Good]	[-]
Microsporum audouinii 9079	[50-100]	[Good]	[Pink-red]
Aspergillus niger 16404	[50-100]	[none to poor]	[-]
Pseudomonas aeruginosa 27853	[50-100]	[none to poor]	[-]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

**Address:-** S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.  
**Email ID:-** biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** [www.biomarklabs.com](http://www.biomarklabs.com)

Rev: January 2026