

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT: Rose Bengal Chloramphenicol Agar Plate**

**PRODUCT CODE NO: PP024-20**

### QUALITY CONTROL TESTS: DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Deep pink coloured medium.
APPEARANCE	Sterile Rose Bengal Chloramphenicol Agar in 85mm disposable plates
CLARITY	Slightly opalescent
STERILITY AFTER 48 HRS	Sterile
PH	7.2 ± 0.2

### QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 22 - 28°C for 48-72 hours.

Organism Description	Inoculum(CFU)	Growth
Bacillus Subtilis 6633	[ $\geq 10^3$ ]	[inhibited]
Escherichia Coli 25922	[ $\geq 10^4$ ]	[inhibited]
Enterococcus Faecalis 19433	[ $\geq 10^4$ ]	[inhibited]
Aspergillus Niger 16404	[50-100]	[luxuriant ]
Saccharomyces cerevisiae 9763 l	[50-100]	[luxuriant ]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

**Address:-** S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.  
**Email ID:-** biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** [www.biomarklabs.com](http://www.biomarklabs.com)

Rev: January 2026