

CERTIFICATE OF ANALYSIS

PRODUCT : Sheep blood agar plate

PRODUCT CODE NO.: PP028-50

QUALITY CONTROL: PHYSICAL

| PARAMETERS | SPECIFICATIONS/ STANDARD VALUES |
|-----------------------|--|
| Colour | Bright Red coloured medium |
| Appearance: | Sterile Sheep Blood Agar in 85mm disposable plates |
| Clarity | Opaque |
| Sterility after 24hrs | Sterile |

QUALITY CONTROL TESTS: MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed after (18-48) hours at (35-37)oC .

| Orgm Desc | Inoculum (CFU) | Growth | Recovery | Haemolysis |
|------------------------------------|----------------|-----------|----------|------------|
| Streptococcus pneumoniae ATCC 6303 | 50-100 | luxuriant | >=70% | alpha |
| Streptococcus pyogenes ATCC 19615 | 50-100 | luxuriant | >=70% | beta |

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.
Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com.

Rev: December 2023