

## CERTIFICATE OF ANALYSIS

**PRODUCT :** Bacillus Cereus Agar Base-MYP

**PRODUCT CODE NO.:** PP035-20

## QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/ STANDARD VALUES
Colour	Light orange to pink
Appearance	Sterile B cereus Agar Base (MYP) in 85mm dis. plates.
Clarity	opaque
Sterility after 24hrs	Sterile

## QUALITY CONTROL TESTS: MICROBIOLOGICAL

## CULTURAL RESPONSE:

Cultural characteristics observed after (18-48 ) hours at (30±2)°C .

Orgm Desc	Inoculum (CFU)	Growth	Colour of colony	Lecithinase Activity
Bacillus cereus ATCC 10876	50-100	Luxuriant	Red	Positive, opaque zone around the colony
Bacillus subtilis (6633)	50-100	Luxuriant	Yellow	Negative
Staphylococcus aureus (25923)	50-100	Luxuriant	Yellow	Positive, opaque zone around the colony
Escherichia coli (25922)	50-100	None-Poor	-	-
Pseudomonas aeruginosa (27853)	50-100	None-Poor	-	-

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.