

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT: Nutrient Agar**

**PRODUCT CODE NO: SB005-100**

### QUALITY CONTROL TESTS: DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Light Amber colored medium
PH	7.4 ±0.2
APPEARANCE	Sterile Nutrient Agar medium in glass bottle.
STERILITY TEST (AFTER 48 HOURS)	Sterile

### QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE:

Cultural characteristics after melting the medium and pouring into sterile petri plates. The plates are inoculated with following test organisms and incubation at 35 -37°C for 18-24 hours.

Organism Description	Inoculum(CFU)	Growth
Escherichia coli 25922	[50-100]	[Good-Luxuriant ]
Staphylococcus aureus 25923	[50-100]	[Good-Luxuriant ]
Pseudomonas aeruginosa 27853	[50-100]	[Good-Luxuriant ]
Streptococcus pyogenes19615	[50-100]	[Good-Luxuriant ]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

**Address:-** S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

**Email ID:-** biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** [www.biomarklabs.com](http://www.biomarklabs.com)

Rev: January 2026