

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT: Sabouraud Dextrose Agar

PRODUCT CODE NO: SB008-100

QUALITY CONTROL TESTS: DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Light Amber coloured medium
PH	5.6 ±0.2
APPEARANCE	Sterile glass bottle containing clear to slightly opalescent Sabouraud Dextrose Agar
STERILITY TEST (AFTER 48 HOURS)	Sterile

QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE:

Cultural characteristics after melting the medium and pouring into sterile petri plates. The plates are inoculated with following test organisms and incubation at 25°C for 7 days.

Organism Description	Inoculum(CFU)	Growth
Candida albicans 10231	[50-100]	[Luxuriant]
Aspergillus niger 16404	[50-100]	[Luxuriant]
Saccharomyces cerevisiae 9763	[50-100]	[Luxuriant]
Trichophyton rubrum 28191	[50-100]	[Luxuriant]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com

Rev: January 2026