

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT: Transport Swabs W/Buffered Sodium Chloride Peptone Solution PH 7.0**

**PRODUCT CODE NO: TS003-50**

### QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Colorless to pale yellow colored medium.
pH	7.3 ± 0.2
Appearance	Sterile Buffered Sodium Chloride Peptone Solution PH 7.0 in tubes with sterile cotton swabs.
Sterility after 48 hrs	Sterile

### QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE

Cultural characteristics observed after 18 –24 hrs. at 35-37°C

Organism Description	Inoculum(CFU)	Recovery
Salmonella typhimuriam (14028)	[50-100]	[Luxuriant]
Escherichia coli 25922	[50-100]	[Luxuriant]
Salmonella abony (NCTC 6017)	[50-100]	[Luxuriant]
Staphylococcus aureus (25923)	[50-100]	[Luxuriant]
Staphylococcus aureus (6538)	[50-100]	[Luxuriant]
Escherichia coli (8739)	[50-100]	[Luxuriant]
Escherichia coli (NCTC9002)	[50-100]	[Luxuriant]
Pseudomonas aeruginosa (27853)	[50-100]	[Luxuriant]
Pseudomonas aeruginosa (9027)	[50-100]	[Luxuriant]
Candida albicans (10231)	[50-100]	[Luxuriant]
Candida albicans (2091)	[50-100]	[Luxuriant]
Bacillus subtilis (6633)	[50-100]	[Luxuriant]
Micrococcus luteus (9341)	[50-100]	[Luxuriant]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

---

**Address:-** S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.  
**Email ID:-** biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** [www.biomarklabs.com](http://www.biomarklabs.com)

Rev: January 2026