

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT: Transport Swabs w/ Cary -Blair Medium

PRODUCT CODE NO: TS004-50

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Light Amber colored medium
pH	7.3 ± 0.2
Appearance	Sterile Cary-Blair Medium in tubes with sterile cotton swabs.
Sterility after 48 hrs	Sterile

QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE

Viability of following organisms was established for a period of 48 hours. Organisms grew luxuriantly when recovered on Tryptone Soya Agar (B039) and incubated at 35 - 37°C for 18-24 hours. Cultural characteristics observed after 18 -24 hrs. at 35-37°C

Organism Description	Inoculum(CFU)	Recovery
Enterobacter aerogenes 13048	[50-100]	[Good-Luxuriant]
Escherichia coli 25922	[50-100]	[Good-Luxuriant]
Klebsiella pneumoniae 13883	[50-100]	[Good-Luxuriant]
Neisseria meningitidis 13090	[50-100]	[Good-Luxuriant]
S. serotype Typhimurium 14028	[50-100]	[Good-Luxuriant]
Shigella flexneri 12022	[50-100]	[Good-Luxuriant]
Vibrio parahaemolyticus 11344	[50-100]	[Good-Luxuriant]
Vibrio cholerae 15748	[50-100]	[Good-Luxuriant]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.
Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, Website:- www.biomarklabs.com

Rev: January 2026