

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT: Transport Swabs w/ Chlamyospore Medium

PRODUCT CODE NO: TS005-50

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Blue colored medium.
pH	5.1 ± 0.2
Appearance	Sterile Chlamyospore Medium in tubes with sterile cotton swabs.
Sterility after 48 hrs	Sterile

QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE

Chlamyospore formation was observed when inoculated and recovered on respective media like Sabouraud Dextrose Agar (B057) and incubated at 22-28⁰C for 48 hours.

Organism Description	Inoculum(CFU)	Growth	Chlamyospores
Candida albicans 10231	[50-100]	[Good-Luxuriant]	[Positive]
Candida krusei 24408	[50-100]	[Good-Luxuriant]	[Negative]
Candida tropicalis 1369	[50-100]	[Good-Luxuriant]	[Negative]
Candida minosa	[50-100]	[Good-Luxuriant]	[Negative]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.
Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com

Rev: January 2026