

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT: Sterile Swabs w/Letheen Broth

PRODUCT CODE NO: TS007-25

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Yellow coloured medium
pH	7.0± 0.2
Appearance	Sterile Swabs w/Letheen Broth in tubes with sterile cotton swabs.
Sterility after 48 hrs	Sterile

QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE

Cultural characteristics observed after an incubation at 35-37°C for 18-48 hours.

Organism Description	Inoculum(CFU)	Growth
Escherichia coli 25922 (00013*)	[50-100]	[Good-Luxuriant]
Staphylococcus aureus subsp. aureus 25923 (00034*)	[50-100]	[Good-Luxuriant]
Escherichia coli 8739 (00012*)	[50-100]	[Good-Luxuriant]
Staphylococcus aureus subsp. aureus 25923 (00032*)	[50-100]	[Good-Luxuriant]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com

Rev: January 2026