

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT: Sterile Swab w/ 0.1% Peptone Water**

**PRODUCT CODE NO: TS008-10**

### QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Clear colorless medium.
pH	7.0± 0.2
Appearance	Sterile 0.1% Peptone water in tube with Sterile Cotton Swabs..
Sterility after 48 hrs	Sterile

### QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE

Viability of following was established for a period of 24 hours. Organisms grew luxuriantly when recovered on Tryptone Soya Agar (B039) and incubates at 35-37°C for 18-24 hours

Organism Description	Inoculum(CFU)	Growth
Escherichia coli 25922	[50-100]	[Luxuriant]
Staphylococcus aureus 25923	[50-100]	[Luxuriant]
Salmonella Typhimurium 14028	[50-100]	[Luxuriant]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

---

**Address:-** S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.  
**Email ID:-** biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** [www.biomarklabs.com](http://www.biomarklabs.com)

Rev: January 2026