

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT: Sterile Swabs w/0.9% Saline & second tube containing SCDM w/BCP

PRODUCT CODE NO: TS010-50

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	0.9% Saline : Colourless clear solution , SCDM w/BCP : Light purple to purple clear solution
pH	7.3± 0.2
Appearance	Each pack contains two tubes ,Sterile Swabs w/0.9% Saline in one tube & second tube containing Sterile Swabs w/ SCDM w/BCP.
Sterility after 48 hrs	Sterile

QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE

Viability of following was established for a period of 24 hours. Organisms grew luxuriantly when recovered on Tryptone Soya Agar (B039) and incubates at 35-37°C for 18-24 hours.

Organism Description	Inoculum(CFU)	Growth
Escherichia coli 25922 (00013*)	[50-100]	[Luxuriant]
Staphylococcus aureus subsp. aureus 25923	[50-100]	[Luxuriant]
Salmonella Typhimurium 14028 (00031*)	[50-100]	[Luxuriant]
Salmonella Abony NCTC6017 (00029*)	[50-100]	[Luxuriant]
Pseudomonas aeruginosa 9027 (00026*)	[50-100]	[Luxuriant]
Streptococcus pneumoniae 6305	[50-100]	[Luxuriant]
Staphylococcus aureus subsp. aureus 6538 (00032*)	[50-100]	[Luxuriant]
Escherichia coli NCTC 9002	[50-100]	[Luxuriant]
Escherichia coli 8739 (00012*)	[50-100]	[Luxuriant]
Bacillus subtilis subsp. spizizenii 6633 (00003*)	[50-100]	[Luxuriant]
# Aspergillus brasiliensis 16404 (00053*)	[50-100]	[Luxuriant]
Candida albicans 10231 (00054*)	[50-100]	[Luxuriant]

Candida albicans 2091 (00055*)	[50-100]	[Luxuriant]
--------------------------------	----------	-------------

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biemark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com

Rev: January 2026