

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT: Transport Swabs w/ Stuart Transport Medium**

**PRODUCT CODE NO: TS011-10**

### QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Whitish to light blue coloured medium
pH	7.4 ± 0.2
Appearance	Sterile Stuart Transport Medium in tubes with sterile viscous swabs
Sterility after 48 hrs	Sterile

### QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE

Viability of following organisms was established for a period of 48hours. Organisms grew luxuriantly when inoculated on Chocolate Agar (B980) and incubated at 35 - 37°C for 18-24 hours

Organism Description	Inoculum(CFU)	Growth
Neisseria gonorrhoeae 19424	[50-100]	[Luxuriant]
Streptococcus pneumoniae 6303	[50-100]	[Luxuriant]
Haemophilus influenzae 35056	[50-100]	[Luxuriant]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

**Address:-** S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

**Email ID:-** biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** [www.biomarklabs.com](http://www.biomarklabs.com)

Rev: January 2026