

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT: Transport Swab w/ Soyabean Casein Digest Medium w/6.5% NaCl**

**PRODUCT CODE NO: TS014-50**

### QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Light yellow colored medium.
pH	7.3 ± 0.2
Appearance	Sterile Soyabean Casein Digest Medium w/6.5% NaCl in tube with sterile swab in individual pack
Sterility after 48 hrs	Sterile

### QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE

Viability of various microorganisms was established for a period upto 24 hours. Organisms grew luxuriantly when inoculated and recovered on Tryptone Soya Agar (B039) and incubated at 35 -37°C for 18-24 hours

Organism Description	Inoculum(CFU)	Growth
Staphylococcus aureus 25923	[50-100]	[Luxuriant]
Staphylococcus aureus subsp. aureus 6538	[50-100]	[Luxuriant]
Staphylococcus epidermidis 12228	[50-100]	[Luxuriant]
Staphylococcus aureus MRSA 43300	[50-100]	[Luxuriant]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

Email ID:- [biomarklabs@gmail.com](mailto:biomarklabs@gmail.com) / [biomarkmktg@gmail.com](mailto:biomarkmktg@gmail.com), Website:- [www.biomarklabs.com](http://www.biomarklabs.com)

Rev: January 2026