

<b>B016</b>	<b>BRILLIANT GREEN BILE AGAR</b>				
<b>Formula</b>					
<b>Ingredients :</b>		<b>gms/lit.</b>			
Peptone		8.25			
Lactose		1.90			
Sodium sulphite		0.205			
Ferric chloride		0.0295			
Monopotassium phosphate		0.0153			
Erioglaucine		0.0649			
Basic fuchsin		0.0776			
Oxgall		0.00295			
Brilliant green		0.0000295			
Agar		10.15			
Final pH (at 25°C) : 6.9 ± 0.2					
<b>Directions :</b>					
Suspend 20.7 grams in 1000 ml distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45-50°C. Mix well and pour into sterile Petri plates. For plating 10 ml quantities of water samples, prepare the medium in double strength.					
<b>Principle :</b>					
Brilliant Green Bile Agar contains peptone as a source of carbon, nitrogen, vitamins and minerals. Lactose is a fermentable carbohydrate. Oxgall (bile) and brilliant green inhibit gram-positive bacteria and most gram-negative bacteria except coliforms. Erioglaucine and basic fuchsin together form the indicator system of the medium. Monopotassium phosphate is a buffering agent. Agar is a solidifying agent. Differentiation of the coliforms is based on fermentation of lactose. When the pH is neutral, colour of the medium is blue while acid production from lactose turns the medium pink and colonies appear pink to deep red depending on the pH change. Colonies of coliform bacteria are deep red surrounded by a pink halo against blue background of the medium. Bacteria that do not ferment lactose form colorless to faint pink colonies. Coliform bacteria typically ferment lactose, producing deep red colonies, while Salmonella spp., which do not ferment lactose, produce colorless to faint pink colonies.					
<b>QC Tests – (I) Dehydrated Medium</b>					
	Colour :	Pink to light purple			
	Appearance :	Homogeneous Free Flowing powder			
<b>(II) Rehydrated medium</b>					
	pH (post autoclaving/heating) :	6.9 ± 0.2			
	Colour (post autoclaving/heating) :	Bluish purple			
	Clarity (post autoclaving/heating) :	Slightly opalescent			
<b>(III) Q.C. Test Microbiological</b>					
Cultural characteristics observed after 18 –24 hours at 35-37°C.					
	MICROORGANISM (ATCC )	GROWTH	COLOUR OF COLONY		
	Enterobacter aerogenes (13048)	Good-luxuriant	Pink		
	Escherichia coli (25922)	Good-luxuriant	Deep red with bile precipitate		
	Salmonella enteritidis (13076 )	Good-luxuriant	Colourless to light pink		
	Staphylococcus aureus (25923)	Inhibited	---		
<b>Precautions :</b>	<ol style="list-style-type: none"> <li>1. For Laboratory Use.</li> <li>2. Follow proper, established laboratory procedures in handling and disposing of infectious materials.</li> <li>3. Possible risk of irreversible effects. Avoid contact with skin and eyes. Do not breathe dust. Wear suitable protective clothing. Keep container tightly closed. Target organ(s): Liver, Thyroid.</li> </ol>				
<b>Limitations :</b>	<ol style="list-style-type: none"> <li>1. Since the nutritional requirements of organisms vary, some strains may be encountered that fail to grow or grow poorly on this medium.</li> <li>2. It is recommended that the medium be prepared just prior to use and if the medium has to be stored, it should be kept in dark. Brilliant Green Bile Agar medium is sensitive to light, particularly direct sunlight. Direct exposure may exhibit a decrease in the productivity of the medium and also the colour of the medium may change from deep blue to purple or red.</li> </ol>				
<b>Use :</b>	For enumeration of coliform bacteria in water, sewage and food.				
<b>Storage :</b>	Dehydrated medium- below 30°C Prepared medium- Between 2 to 8°C.				
<b>Packing :</b>	500 gm. bottle				
<b>Product profile:</b>	Reconstitution	Quantity on Preparation (500g)	pH (25°C)	Supplement	Sterilization
<b>B016</b>	20.7g/l	24.154L	6.9 ± 0.2	NIL	121°C / 15 minutes

**Disclaimer:**

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications. The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.