

**BIOMARK Laboratories-INDIA**

[www.biomarklabs.com](http://www.biomarklabs.com)

**TECHNICAL SHEET**

<b>B039</b>	<b>SOYABEAN CASEIN DIGEST AGAR (TRYPTONE SOYA AGAR)</b>		
<b>Formula</b>			
<b>Ingredients:</b>	<b>gms/lit.</b>		
Tryptone	15.00		
Soya peptone	5.00		
Sodium chloride	5.00		
Agar	15.00		
Final pH (at 25°C):7.3 + 0.2			
<b>Directions:</b>			
Suspend 40 grams in 1000 ml distilled/purified water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. If desired, aseptically add 5% v/v defibrinated blood in previously cooled medium to 45-50°C for cultivation. Mix well and pour into sterile Petri plates.			
<b>Principle:</b>			
The medium with addition of blood provides perfectly defined hemolysis zones, while preventing the lysis of erythrocytes due to its sodium chloride content.It's simple and inhibitor-free composition makes it suitable for the detection of antimicrobial agents in the food and other products.The combination of tryptone and soya peptone makes this media nutritious by providing amino acids and long chain peptides for the growth of microorganisms. Sodium chloride maintains the osmotic balance.			
<b>Type of specimen :</b> Pharmaceutical samples, Clinical samples- urine, faeces, abscess etc.			
<b>Specimen Collection and Handling:</b>			
For clinical samples follow appropriate techniques for handling specimens as per established and current guidelines of clinical microbiology.			
For Pharmaceutical samples, follow appropriate techniques for sample collection,handling and processing as per standard and current guidelines of pharmacopeias.			
After use, contaminated materials must be sterilized by autoclaving before discarding.			
<b>QC Tests – (I)Dehydrated Medium</b>			
Colour:	Cream to light yellow		
Appearance:	Homogeneous Free Flowing powder		
<b>(II)Rehydrated medium</b>			
pH (post autoclaving/heating):	7.3 ± 0.2		
Colour (post autoclaving/heating):	a) Light yellowb) After addition of5-7%w/v sterile defibrinated blood : Cherry red		
Clarity (post autoclaving/heating):	a) Clear slightly opalescentb) Opaque		
<b>(III)Q.C. Test Microbiological</b>			
Cultural characteristics was observed after an incubation for Bacterial at 30-35°C 18-24 hours and for Fungal at 30-35°C <=5days.			
MICROORGANISM (ATCC )	GROWTH W/ BLOOD	HAEMOLYSIS	
Bacillus subtilis (6633)	Luxuriant	None	
Staphylococcus aureus (6538)	Luxuriant	β	
Staphylococcus aureus (25923)	Luxuriant	β	
Escherichia coli (25922)	Luxuriant	None	
Escherichia coli (8739)	Luxuriant	None	
Escherichia coli (NCTC9002)	Luxuriant	None	
Escherichia coli (11775)	Luxuriant	None	
Escherichia coli (NCTC13167)	Luxuriant	None	
Pseudomonas aeruginosa (10145)	Luxuriant	None	
Pseudomonas aeruginosa (27853)	Luxuriant	None	
Pseudomonas aeruginosa (9027)	Luxuriant	None	
Salmonella Abony (NCTC6017)	Luxuriant	None	

Refer disclaimer Overleaf

Page 01 of 02

Rev:July 2022

**BIOMARK Laboratories-INDIA**

[www.biomarklabs.com](http://www.biomarklabs.com)

**TECHNICAL SHEET**

	Micrococcus luteus (9341)	Luxuriant	None	
	Streptococcus pneumoniae (6305)	Luxuriant	None	
	Salmonella Typhimurium (14028)	Luxuriant	None	
	Enterococcus faecalis (29212)	Luxuriant	None	
	Candida albicans (10231)	Luxuriant	None	
	Candida albicans (2091)	Luxuriant	None	
	Clostridium perfringenes (13124)	Luxuriant	None	
	Aspergillus brasiliensis (16404)	Good-Luxuriant	None	
<b>Warning &amp; Precautions :</b>	1. For In vitro diagnostic Use.By professionals only.			
	2. Read the label carefully before opening the container.Wear PPE wares.Follow established good microbiology laboratory practices while handling specimens and cultures and take standard precautions for handling clinical specimens.			
	3. For safety guidelines refer individual safety data sheet.			
<b>Limitations :</b>	1. Since the nutritional requirements of organisms vary, some strains may be encountered that fail to grow or grow poorly on this medium.			
	2.This medium is general purpose medium and may not support the growth of fastidious organisms.			
	3. Further biochemical and serological tests must be carried out for confirmation.			
<b>Use:</b>	It is a general-purpose medium used for cultivation of a wide variety of microorganisms from clinical and non-clinical samples and for sterility testing in pharmaceutical procedures.			
<b>Storage:</b>	Dehydrated medium- below 30°C Prepared medium– Between 2 to 8°C.			
<b>Disposal:</b>	Ensure safe disposal by autoclaving/or incineration of used or usable preparation of this product. Follow established laboratory procedures while disposing all infectious material and those coming in contact must be decontaminated and disposed off with existing laboratory technics.			
<b>Packing:</b>	500 gm. bottle			
<b>Product profile:</b>	Reconstitution	Quantity on Preparation (500g)	pH (25°C)	Supplement
<b>B039</b>	40g/l	12.5L	7.3 ± 0.2	5% v/v defibrinated blood
				121°C / 15 minutes

**Disclaimer:**

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.