

TECHNICAL SHEET

B155	CYSTINE TRYPTONE AGAR		
Formula			
Ingredients :	gms/lit.		
Casein enzymic hydrolysate	20.00		
L-Cystine	0.50		
Sodium chloride	5.00		
Sodium sulphite	0.50		
Phenol red	0.017		
Agar	2.50		
Final pH (at 25°C) : 7.3 ± 0.2			
Directions :			
Suspend 28.51 grams in 1000 ml distilled water. Heat to boiling to dissolve the medium completely. Dispense in tubes in 8-10 ml amounts. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 50°C and add appropriate carbohydrate (0.5 to 1.0% if desired). Mix well and allow the tubed medium to cool in an upright position.			
Principle :			
Casein enzymic hydrolysate, L-cystine supplies the nutrients necessary to support the growth of fastidious microorganism. Carbohydrate fermentation is detected by a visible colour change of the medium due to the incorporation of the pH indicator dye, phenol red. When the organism metabolizes the carbohydrate present, organic acids are produced and the medium becomes acidified. However, the peptones present in the medium are also degraded by the bacteria present and yield substances that are alkaline in pH. The phenol red indicator changes from reddish-orange to yellow when the amount of acid produced by carbohydrate fermentation is greater than the alkaline end products of the peptone degradation. The colour change with phenol red occurs around pH 6.8, near the original pH of the medium.			
QC Tests – (I)Dehydrated Medium			
Colour :	Light yellow to light pink		
Appearance :	Homogeneous Free Flowing powder		
(II)Rehydrated medium			
pH (post autoclaving/heating) :	7.3 ± 0.2		
Colour (post autoclaving/heating) :	Orange to red		
Clarity (post autoclaving/heating) :	Clear to slightly opalescent		
(III)Q.C. Test Microbiological			
Cultural characteristics observed after an incubation at 35-37°C for 4-18 hours or longer if necessary.			
MICROORGANISM (ATCC)	GROWTH	MOTILITY	ACID IN PRESENCE OF DEXTROSE
Escherichia coli (25922)	Good – luxuriant	+	Positive reaction, yellow colour
Neisseria meningitidis (13090)	Good	-	Positive reaction, yellow colour
Neisseria gonorrhoeae (19424)	Good	-	Positive reaction, yellow colour
Streptococcus pneumoniae (6303)	Good	-	Positive reaction, yellow colour
Key: For motility + =positive, growth away from stabline causing turbidity for motility - = negative, growth along the stabline, surrounding medium remains clear			

Refer disclaimer Overleaf

TECHNICAL SHEET

Precautions :	1. For Laboratory Use. 2. Follow proper, established laboratory procedures in handling and disposing of infectious materials.				
Limitations :	1. Since the nutritional requirements of organisms vary, some strains may be encountered that fail to grow or grow poorly on this medium. 2. CTA requires a heavy inoculum. 3. Prolonged incubation may lead to changes in pH indicator or abnormal lactose / sucrose reactions with Neisseria pathogens. 4. Neisseria species usually produce acid only in the area of stabs (upper third). If there is a strong acid (yellow color) throughout the medium, a contaminating organism may be present. If in doubt about a tube containing a Neisseria species, a Gram stain and oxidase test should be performed on the growth.				
Use :	For maintenance, subculturing, detection of motility etc. With added carbohydrates, it can be also used for fermentation reactions of fastidious organisms.				
Storage :	Dehydrated medium- below 30°C Prepared medium- Between 2 to 8°C.				
Packing :	500 gm. Bottle				
Product profile:	Reconstitution	Quantity on Preparation (500g)	pH (25°C)	Supplement	Sterilization
B155	28.51g/l	17.537 L	7.3 ± 0.2	nil	121°C / 15 minutes

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.