BIOMARK Laboratories-INDIA

www.biomarklabs.com

TECHNICAL SHEET

B333	THAYER MARTIN MEDIUM BASE				
Formula					
Ingredients:	gms/lit.				
Peptone, speci	ial 23.00				
Starch	1.00				
Sodium chloric	de 5.00				
Agar	13.00				
Final pH (at 25	5°C): 7.0 <u>+</u> 0.2				
Directions:	_				
Cuspend 21 0	grams in 4E0 ml distilled water. Heat to beiling to dissolve the medium completely. Sterilize h				

Suspend 21.0 grams in 450 ml distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45°C. Aseptically add 50 ml of sterile lysed blood and rehydrated contents of one vial of Vitamino Growth Supplement (BF037) and V.C.N Supplement (BF059) or V.C.N.T Supplement (BF60). If desired GC Supplement with Antibiotics (BF058) can be used as a single supplement. Mix well before pouring into sterile Petri plates. If Hemoglobin (BF036) is used suspend 21.0 grams of Thayer Martin Medium Base in 250 ml distilled water. Heat to boiling to dissolve the medium completely. Prepare 250 ml of 2% hemoglobin solution. Sterilize separately by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45°C. Mix both and add the supplements as above.

Principle:

Special peptone provides nutrients to the organisms while starch neutralizes the toxic fatty acids if present in the agar. Haemoglobin provides the X factor whereas the V factor (N.A.D.) is provided by the added supplement. Supplement (BF037) also supplies vitamins, amino acids, coenzymes etc. which enhances the growth of pathogenic Neisseria. Vancomycin and colistin inhibits gram-positive and gram-negative bacteria respectively. Nystatin inhibits fungi. This medium may inhibit Haemophilus species.

liesbectively. In	ystatiii iiiiiibits iu	ngı. Tilis medi	uiii iiiay	y IIIIIIDIL	naemophilus species.			
QC Tests - (I)Dehydrated Medium					·			
Colour:	Colour:			Cream to yellow				
Appearance:			Homogeneous Free Flowing powder					
(II)Rehydrated medium								
pH (post autoclaving/heating):				7.0 ± 0.2				
Colour (post autoclaving/heating):			a) Basal medium : Yellow					
				b) With addition of haemoglobin and supplements : Chocolate				
Clarity (post autoclaving/heating):			a) Clear to slightly opalescent					
			b) Opaque					
	: Microbiologica							
Cultural characteristics observed with added sterile lysed blood/Haemoglobin solution (BF036), Vitamino								
	Growth Supplement (BF037) and V.C.N. Supplement (BF059)/V.C.N.T. Supplement (BF060) after an							
	tion at 35-37°C for 18-48 hours.							
MICROORGANISM (ATCC) GROWTH Neisseria meningitides (13090) Good – luxi				COLOUR OF COLONY				
Neisseria gonorrhoeae (19424) Good - lux			kuriant	uriant small, grayishwhite to colourless, mucoid				
Escherichia coli (25922) Inhibited								
	irabilis (25933)	Inhibited	-					
Precautions: 1. For Laboratory Use.								
				aboratory procedures in handling and disposing of infectious				
	materials.							
Limitations :	1 7,							
	that fail to grow or grow poorly on this medium.							
Use:	For selective isolation of Gonococci from pathological specimens.							
Storage:	Dehydrated medium- below 30°C Prepared medium- Between 2 to 8°C.							
Packing:	500 gm. bottle Reconstitution Quantity on pH (25°C) Supplement Steriliza							
Product	Reconstitution	Quantity on	pπ	1 (25°C)	Supplement	Sterilization		
profile:		Preparation (500g)						
B333	42g/l	11.904L	7.0	0 ± 0.2	Sterile lysed blood/BF036 & BF037	121°C /15		
					& BF059 or BF060	min.		

Refer disclaimer Overleaf

Page 01 of 02

BIOMARK Laboratories-INDIA www.biomarklabs.com

TECHNICAL SHEET

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

Page 02 of 02

Rev: January 2025