

TECHNICAL SHEET

BA042

PAPANICOLAU-EA-36

Formula

Reagent:

Light green SF yellowish (0.5% alcoholic solution)	4.50 ml
Bismark brown (0.5% alcoholic solution)	10.00 ml
Eosin Y (0.5% alcoholic solution)	45.00 ml
Phosphotungstic acid	200.00 mg
Li2CO3 (Saturated aqueous solution)	1 drop

Quality control:

Appearance: Dirty green coloured solution

Directions:

1. Fix smear immediately in 95% alcohol for 5-15 minutes.
2. Rinse in 70% alcohol, 50% alcohol and distilled water.
3. Stain in Harris haematoxylin (without acetic acid) for 5-10 minutes.
4. Rinse in distilled water.
5. Rinse 3-4 times in 0.5% aqueous solution of HCL. Rinse thoroughly with water.
6. Leave for 1 minute in a weak solution of lithium carbonate. Rinse thoroughly in water.
7. Rinse in distilled water, 50% alcohol, 70% alcohol, 80% alcohol and 95 % alcohol.
8. Stain in the papanicolau-OG-6(BA041) for 1 minute.
9. Rinse 5-10 times in 95 % alcohol.
10. Stain in the papanicolau-EA -36(BA042) for 2 minutes.
11. Rinse 5-10 times in 95 % alcohol.
12. Rinse in absolute alcohol, then in a mixture of equal parts of absolute alcohol and xylene and then in xylene.
13. Mount in neutral medium and observe under microscope.

Quality Control:

Microscopy: Nuclei: Blue Cytoplasm: Grey or pale blue

Storage: Below 30°C.

Packing: 100 ml

Warning & Precautions :

In Vitro diagnostic Use only. For laboratory use. Read the label before opening the container. Wear protective gloves/protective clothing/ eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Use: Staining of malignant cells

Disposal: User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques.

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.