

**BA092 Potassium permanganate**

**Formula**

**Reagent:**

Potassium Permanganate	0.100 g
Distilled water	100.000 ml

**Directions:**

- 1) Thoroughly flame one side of a clean glass slide to be used for the smear and allow to cool before smearing.
- 2) Prepare a smear from the specimen on the flamed side of the slide and allow to air dry.
- 3) Fix the smear by passing the slide through a flame.
- 4) Flood the slide with phenolic auramine(BA090) Heat to steaming for 10-15 minutes with a low flame; do not boil the stain and do not permit drying of the smear.
- 5) Rinse with distilled water and drain off excess liquid.
- 6) Flood the slide with decolorizer (BA091) for 2-3 min. Slide will still appear pink.
- 7) Rinse thoroughly with distilled water, drain off excess.
- 8) Flood slide with counterstain (BA092) for 3-4 min. Do not allow slide to dry.
- 9) Rinse thoroughly with distilled water and allow to air dry.
- 10) Examine the slide by fluorescent microscope.

**Quality control:**

**Appearance:** Dark purple coloured solution

**Microscopy:** Acid Fast Bacilli: luminous yellow rods in dark field.

**Storage:** Below 30°C in tightly closed container and away from bright light

**Packing:**100 ml

**Warnig & precaution:**

In Vitro diagnostic Use only. For Laboratory use. Read the label before opening the container. Wear protective gloves/protective clothing/ eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets

**USE:** It is used as counter stain for the acid-fast bacteria or other cellular matter in staining by fluorescence microscopy.

**Disposal:**

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques.

Refer disclaimer Overleaf

**BIOMARK Laboratories-INDIA**

[www.biomarklabs.com](http://www.biomarklabs.com)

**TECHNICAL SHEET**

**Disclaimer:**

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.