

**BA101 GORDEN-McLEOD REAGENT (OXIDASE REAGENT)**

**Formula**

**Reagents:**

N, N-Dimethyl p-phenylene diamine hydrochloride 1.50 g  
 Distilled water 100.00 ml

**Directions:**

1. Place 2-3 drops of reagent on to a filter paper in a petri dish.
2. With a platinum wire loop (not nichrome), plastic loop or glass rod, smear some of the colonies from Soyabean casein digest agar (B039) on the prepared filter paper.
3. Observe for appearance of deep blue - purple color within 10 seconds.

**Quality Control:**

**Appearance:** Black colored clear solution

**Microbiological testing:**

<b>MICROORGANISM (ATCC)</b>	<b>OXIDASE TEST</b>
Neisseria gonorrhoeae (19424)	Positive (development of purple-blue colour)
Pseudomonas aeruginosa (27853)	Positive (development of purple-blue colour)
Staphylococcus aureus (25923)	Negative (No colour change)

**Storage:** Between 10°C - 30 °C.

**Packing:** 100 ml bottle

**Warning & Precautions :**

In Vitro diagnostic Use only. For professional use only. For Laboratory Use. Read the label before opening the container. Wear protective gloves/protective clothing/ eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets. \*Hazardous chemical.Handle carefully.

**Use:** Oxidase reagent is used to carry out the oxidase test, to determine the presence of oxidase enzymes

**Disposal:**

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques.

**Disclaimer:**

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.