

BA125 BUFFER SOLUTION pH 4.0

Formula

Preparation:

Citric acid	5.90g	
Disodium hydrogen phosphate		5.50 g
Distilled water		1000.00 ml

Precautions :

1. For Laboratory Use.

Use: pH meter standardization

Quality Control:

Appearance: Colourless clear liquid

pH: 4.00 ± 0.05

Storage: Below 30°C

Packing :1000 ml bottle

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.