BIOMARK Laboratories-INDIA

www.biomarklabs.com

TECHNICAL SHEET

PP016 Mueller Hinton Aga	r Plate
Formula	
Ingredients:	gms/lit.
Meat, infusion from#	300.00
Casein acid hydrolysate	17.50
Starch	1.50
Agar	17.0
Final pH (at 25°C): 7.3 <u>+</u> 0.2	
Directions:	
Label the ready to use plate (PP016).Either streak, inoculate or surface spread the
test inoculum (50-100 CFU) aseptica	lly on the plate.
Principle:	hydrolysate provide nitrogenous compounds,
substances present in the medium influence the disc diffusion suscept depth, disc potency, medium pH organisms. A standardized suspensi surfaceof the medium. Paper di antimicrobial agents are then place zones of inhibition around each disc by comparingwith CLSI standards. M by disc diffusion method with slow of	nutrients.Starch is added to absorb any toxic .Agar is the solidifying agent.Different factors otibility tests as, inoculum concentration,agar and beta – lactamase production by test on of the organism is swabbed over the entire scs impregnated with specific amounts of d on the surface of themedium, incubated and are measured. The susceptibility is determined lueller Hinton Agar is not appropriate for assay growing organisms, anaerobes and capnophiles. reased incubation may cause deterioration of ecise readings.
pH:	7.3 ± 0.2
Color:	Light Amber coloured medium
Appearance:	Sterile Mueller Hinton Agar in 85mm disposable
	plates.
(II)Sterility test	Passes release criteria
(III)Q.C. Test Microbiological	
	after incubation at 35-37°C for 18-24 hours.
MICROORGANISM (ATCC)	GROWTH
Escherichia coli 25922	luxuriant
Staphylococcus aureus25923	luxuriant
Pseudomonas aeruginosa27853	luxuriant
Enterococcus faecalis 29212	luxuriant

Refer disclaimer Overleaf

Page 01 of 02

BIOMARK Laboratories-INDIA www.biomarklabs.com TECHNICAL SHEET

Precautions :	1. In Vitro diagnostic use only.	
	2. Read the label before opening the container	
Limitations :	: 1. Since the nutritional requirements of organisms vary, some strains may be	
	encountered that fail to grow or grow poorly on this medium.	
Use:	For the determination of susceptibility microorganisms to antimicrobial agents.	
Storage:	Store between 15-25°C. Use before expiry date on the label.	
Packing:	20/50 disposable plates.	

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal ortherapeutic use but for laboratory, diagnostic, research orfurther manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

Page 02 of 02