

**TECHNICAL SHEET**

<b>PP017</b>	<b>Mueller Hinton Agar Plate w/ 2% NaCl</b>	
<b>Formula</b>		
<b>Ingredients:</b>	<b>gms/lit.</b>	
Meat, infusion from#	300.00	
Casein acid hydrolysate	17.50	
Starch	1.50	
Sodium Chloride	20.00	
Agar	17.00	
Final pH (at 25°C): 7.3± 0.2		
<b>Directions:</b>		
Label the ready to use plate (PP017). Either streak, inoculate or surface spread the test inoculum (50-100 CFU) aseptically on the plate.		
<b>Principle:</b>		
Mueller Hinton Agar contains beef extract, acid hydrolysate of casein, and soluble starch which are necessary for bacterial growth. Starch is a protective colloid against toxic materials present in the medium. Sodium chloride supplies essential electrolytes and enhances the growth of staphylococci. The gradient method is based on the concepts of both the dilution and diffusion tests, but differs from conventional disk methods by the use of a preformed and stable antibiotic gradient. When the gradient is applied to the inoculated agar plate, there is an immediate release of the antibiotic into the agar matrix. A continuous and exponential gradient of antibiotic concentration is created underneath the carrier, and after incubation a symmetrical inhibition ellipse centered along the carrier is seen. performance can be documented according to CLSI guidelines		
<b>(I) QC Tests</b>		
pH:	7.3 ± 0.2	
Color:	Light Amber coloured medium	
Appearance:	Sterile Mueller Hinton Agar w/ 2% NaCl in 85mm disposable plates.	
<b>(II) Sterility test</b>	Passes release criteria	
<b>(III) Q.C. Test Microbiological</b>		
	Cultural characteristics observed after incubation at 35-37°C for 18-24 hours. performance can be documented according to CLSI guidelines	
MICROORGANISM (ATCC)	GROWTH	
Staphylococcus aureus 29213	luxuriant	
Staphylococcus aureus 43300	luxuriant	

Refer disclaimer Overleaf

TECHNICAL SHEET

<b>Precautions :</b>	1. In Vitro diagnostic use only. 2. Read the label before opening the container
<b>Limitations :</b>	1. Since the nutritional requirements of organisms vary, some strains may be encountered that fail to grow or grow poorly on this medium.
<b>Use:</b>	For the quantitative determination of susceptibility to methicillin and oxacillin when testing Staphylococci directly from colonies grown on nonselective media.
<b>Storage:</b>	Store between 2-8°C. Use before expiry date on the label.
<b>Packing:</b>	20/50 disposable plates.

**Disclaimer:**

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.