

TECHNICAL SHEET

PP022	Pseudomonas Pyocyanin Agar Plate	
Formula		
Ingredients:	gms/lit.	
Peptic digest of animal tissue	20.00	
Potassium sulphate	10.00	
Magnesium chloride	1.40	
Agar	15.00	
Final pH (at 25°C): 7.0 ± 0.2		
Directions:		
Label the ready to use plate (PP022). Either streak, inoculate or surface spread the test inoculum (50-100 CFU) aseptically on the plate.		
Principle:		
Peptone provides the carbon and nitrogen sources required for good growth. Glycerol is a carbon source. Magnesium chloride and potassium sulfate stimulate pyocyanin production. Agar is a solidifying agent. This medium enhances the elaboration of pyocyanin but inhibits the formation of fluorescein pigment. The fluorescein pigment diffuses from the colonies of Pseudomonas into the agar and shows blue coloration. Some Pseudomonas strains produce small amounts of fluorescein resulting in a blue-green coloration. Potassium sulphate and magnesium chloride, which enhances the pyocyanin production and suppresses the fluorescein production. A pyocyanin-producing Pseudomonas strain will usually also produce fluorescein. It must, therefore, be differentiated from other simple fluorescent pseudomonads by other means. Temperature can be a determining factor as most other fluorescent strains will not grow at 35°C. Rather, they grow at 25-30°C.		
(I) QC Tests		
pH:	7.0 ± 0.2	
Color:	Yellow coloured medium.	
Appearance:	Sterile Pseudomonas pyocyanin Agar in 85mm disposable plates.	
(II) Sterility test		
Passes release criteria		
(III) Q.C. Test Microbiological		
Cultural characteristics observed after incubation at 35-37°C for 18-24 hours.		
MICROORGANISM (ATCC)	GROWTH	COLOR OF COLONY
Pseudomonas aeruginosa 17934	luxuriant	Blue green
Pseudomonas aeruginosa 9027	luxuriant	Blue green

Refer disclaimer Overleaf

TECHNICAL SHEET

Precautions :	1. In Vitro diagnostic use only. 2. Read the label before opening the container
Limitations :	1. Since the nutritional requirements of organisms vary, some strains may be encountered that fail to grow or grow poorly on this medium.
Use:	For detection of pyocyanin production by Pseudomonas species
Storage:	Store between 15-25°C. Use before expiry date on the label.
Packing:	20/50 disposable plates.

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.