BIOMARK Laboratories-INDIA

www.biomarklabs.com

TECHNICAL SHEET

Transport Swabs w/ Chlamydospore Medium

Directions:
Using the packed swab, provided along with the media containing tube, collect the
sample to be transported. insert the capped swab with the sample till the bottom of
the medium Tighten the cap firmly. The specimen will be preserved during
transportation and also the viability of the organisms will be maintained but it will
diminish over the time. Some growth of contaminants may occur during longer
period of transport. After the transportation, the specimen should be inoculated in
proper medium as soon as possible. The cultures on transport swabs must not be

kept at room temperature for more than 24 hours.

Principle:

TS005

Ammonium sulphate acts as sources of ions that simulate metabolism. Monopotassium phosphate provides buffering to the medium. Biotin provides the necessary vitamins required for metabolism. Purified polysaccharide acts as a source of carbon. Trypan blue is a vital dye absorbed selectively by the chlamydospores and imparts blue color to chlamydospores, whereas the filaments are colorless. The medium is chemically defined, semisolid supporting the survival of Candida species for a sufficiently long period. Sterile cotton swabs allow absorption of specimen material while polystyrene shaft allows semi flexibility to the swab stick, aiding in collection.

collection.				
(I) QC Tests				
pH:		5.1±0.2		
Color:	Color: Blue colored medium.			
Appear	ance:	Sterile Chlamydospore Medium in tubes with sterile		
		cotton swabs.		
(II)Sterility test		Passes release criteria		
(III)Q.C. Test Microbiological				
Chlamydospore formation was observed when inoculated and recovered on				
respective media like Sabouraud Dextrose Agar (B057) and incubated at 22-28°C				
for 48 hours.				
MICROO	RGANISM (ATCC)	GROWTH	CHLAMYDOSPORES	
Candida	a albicans 10231	good-luxuriant	Positive	
Candida	a kruisei 24408	good-luxuriant	Negative	
Candida	a tropicalis 1369	good-luxuriant	Negative	
Candida	a minosa	good-luxuriant	Negative	

Refer disclaimer Overleaf

Page 01 of 02

Rev: January2022

BIOMARK Laboratories-INDIA www.biomarklabs.com TECHNICAL SHEET

Precautions:	1. In Vitro diagnostic use only.		
	2. Read the label before opening the container		
Limitations :	1. Since the nutritional requirements of organisms vary, some strains may be		
	encountered that fail to grow or grow poorly on this medium.		
Use:	Recommended for recovery of Candida albicans from clinical specimens.		
Storage:	Store between 2-8°C with caps firmly tighten. DO NOT FREEZE. Use before expiry		
	date on label.		
Packing:	7ml of medium in 50 tubes with sterile swabs provided in individual pack.		

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

Page 02 of 02

Rev: January2022